Test report number: ISO 202111\_03295A\_engl

### **Test Report**

## Toxicity testing of a material extract, Cytotoxicity Test

Test report number: ISO 202111-03295A\_engl

commissioned by:

PerformCoat Europe AG Widenstrasse 7

CH-9464 Rüthi / SG, Switzerland

Test materials:



CYTOX biological testing of medical devices Gottlieb-Keim-Straße 60 D-95448 Bayreuth, Germany

tel. +49-921-1511-254 fax +49-921-1511-255 mobile +49-179-5102577

info@cytox.de www.cytox.de

Dec 06th 2021

# Stainless steel test specimens coated with titanium nitride (TiN) by PVD process

Test material received: Nov 19<sup>th</sup> 2021 Test performed: Dec 03<sup>rd</sup> 2021

Result: The stainless steel test specimens coated with titanium nitride

(TiN) by PVD process did not cause a cytotoxic effect.

# Description of the test procedure:


Normative References: ISO 10993-5 (2009); ISO 10993-1 (2018); ISO 10993-12 (2021)

The materials were extracted for 24 h at 37°C and a partial presssure of 5 % carbon dioxide in extraction medium (DMEM medium with antibiotics, without fetal calf serum [FCS]). The material surface / extraction volume ratio was 3 cm² per ml extraction medium. After extraction the extraction medium was sterile filtered and supplemented with sterile FCS (concentration of FCS in extraction medium: 10 %). The FCS-supplemented extraction medium was pipetted under sterile conditions on precultivated cells of the mouse fibroblastic cell line L929 and incubated for 48 h at 37°C and a partial pressure of 5 % carbon dioxide. The extract was tested in four dilutions (90 %, 30 %, 10 % and 3,3 %). Each dilution was tested in four parallel experiments.

Tergitol<sup>TM</sup> 1% was used as a toxic control substance (concentration in the experiment: 1 % v/v). Cell culture medium was used as a non-toxic control. After the 48 h incubation period the protein content of the cell culture was determined by the method according to Bradford.

CYTOX biological testing of medical devices

#### Results:



	Protein content n=4					
	Tergitol ™1	%Culture medium	Extract 90 %	Extract 30 %	Extract 10 %	Extract 3,3 %
Mean	2,76	49,07	47,89	50,11	50,03	48,55
Standarddev.	0,89	0,69	0,55	0,71	0,79	1,18

In the presence of Tergitol $^{\text{TM}}$  1% in the cell culture medium 5,6 % of the protein content compared to the negative control was reached. This value is within the valid range of 15 % protein content or less compared to the negative control.

Materials are evaluated as cytotoxic, if the material extract leads to a protein content of the test cells of less than 70 % compared to the negative control. This was not the case in this test. The material extract did not show a cytotoxic effect.

Test report number: ISO 202111\_03295A\_engl

Result: The stainless steel test specimens coated with titanium nitride (TiN) by PVD process did not cause a cytotoxic effect.

**Explanatory notes:** 

none

Test performed by: Datam Scholoten

authorized by: Vistam/Scholdin / CEO CYTOX)

It is not allowed to publish only parts of this test report without written approval of CYTOX.